

**बिड दस्तावेज़ / Bid Document**

| <b>बिड विवरण/Bid Details</b>   |   |
|--|---|
| <b>बिड बंद होने की तारीख/समय /Bid End Date/Time</b>  | 16-01-2026 18:00:00   |
| <b>बिड खुलने की तारीख/समय /Bid Opening Date/Time</b>   | 16-01-2026 18:30:00   |
| <b>बिड पेशकश वैधता (बंद होने की तारीख से)/Bid Offer Validity (From End Date)</b>   | 180 (Days)  |
| <b>मंत्रालय/राज्य का नाम/Ministry/State Name</b>   | Ministry Of Education   |
| <b>विभाग का नाम/Department Name</b>  | Department Of Higher Education  |
| <b>संगठन का नाम/Organisation Name</b>  | Delhi University  |
| <b>कार्यालय का नाम/Office Name</b>   | Vallabhbhai Patel Chest Institute   |
| <b>कुल मात्रा/Total Quantity</b>   | 1   |
| <b>वस्तु/श्रेणी /Item Category</b>   | Real Time PCR Machine (V2) (Q2) ( PAC Only )  |
| <b>टर्नओवर के लिए एमएसई को छूट प्राप्त है / MSE Relaxation for Turnover</b>  | Yes   Complete  |
| <b>स्टार्टअप के लिए अनुभव के बर्षों और टर्नओवर से छूट प्रदान की गई है /Startup Relaxation for Years of Experience and Turnover</b>   | No  |
| <b>विक्रेता से मांगे गए दस्तावेज़/Document required from seller</b>  | <p>Certificate (Requested in ATC),Additional Doc 1 (Requested in ATC)</p> <p>*In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer</p> |
| <b>क्या आप निविद्याकर्ता द्वारा अपलोड किए गए दस्तावेज़ों को निविद्या में भाग लेने वाले सभी निविद्याकर्ता को दिखाना चाहते हैं? संदर्भ मेनू है/Do you want to show documents uploaded by bidders to all bidders participated in bid?</b> | <p>Yes (Documents submitted as part of a clarification or representation during the tender/bid process will also be displayed to other participated bidders after log in)</p>   |
| <b>बिड लगाने की समय सीमा स्वतः नहीं बढ़ने के लिए आवश्यक बिड की संख्या / Minimum number of bids required to disable automatic bid extension</b>   | 1   |
| <b>दिनों की संख्या, जिनके लिए बिड लगाने की समय-सीमा बढ़ाई जाएगी / Number of days for which Bid would be auto-extended</b>  | 10  |

| बिड विवरण/Bid Details  |                             |
|--|-----------------------------|
| ऑटो एक्सटेंशन अधिकतम कितनी बार किया जाना है। /<br><b>Number of Auto Extension count</b>  | 3                           |
| बिड से रिवर्स नीलामी सक्रिय किया/Bid to RA enabled   | No                          |
| बिड का प्रकार/Type of Bid  | Two Packet Bid              |
| तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय /Time allowed for Technical Clarifications during technical evaluation  | 5 Days                      |
| निरीक्षण आवश्यक (सूचीबद्ध निरीक्षण प्राचिकरण /जेम के साथ पूर्ण पंजीकृत एजेंसियों द्वारा)/Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM) | No                          |
| मूल्यांकन पद्धति/Evaluation Method   | Total value wise evaluation |
| वित्तीय दस्तावेज की आवश्यकता है / Financial Document Required  | Yes                         |
| मध्यस्थता खंड/Arbitration Clause   | No                          |
| सुलह खंड/Mediation Clause  | No                          |

#### ईएमडी विवरण/EMD Detail

|                             |                     |
|-----------------------------|---------------------|
| एडवाइजरी बैंक/Advisory Bank | State Bank of India |
| ईएमडी राशि/EMD Amount       | 62500               |

#### ईपीबीजी विवरण /ePBG Detail

|  |                     |
|--|---------------------|
| एडवाइजरी बैंक/Advisory Bank                                      | State Bank of India |
| ईपीबीजी प्रतिशत (%) /ePBG Percentage(%)                          | 5.00                |
| ईपीबीजी की आवश्यक अवधि (माह) /Duration of ePBG required (Months) | 62                  |

(a). जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित केटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज प्रस्तुत करने हैं। एमएसई केटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।/EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b). ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए। / EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

#### लाभार्थी /Beneficiary :

Director

Vallabhbhai Patel Chest Institute, Department of Higher Education, Delhi University, Ministry of Education

(Director, VPCI)

बोली विभाजन लागू नहीं किया गया/ Bid splitting not applied.

#### एमआईआई खरीद वरीयता/MII Purchase Preference

|  |    |
|--|----|
| एमआईआई खरीद वरीयता/MII Purchase Preference | No |
|--|----|

#### एमआईआई के लिए सक्षम प्राधिकारी का विवरण:/Details of the Competent Authority for MII

|   |   |
|---|---|
| सक्षम प्राधिकारी का नाम/Name of Competent Authority   | Prof. Raj Kumar   |
| सक्षम प्राधिकारी का पदनाम/Designation of Competent Authority  | Director, VPCI  |
| सक्षम प्राधिकारी का कार्यालय/विभाग/प्रभाग/Office / Department / Division of Competent Authority                                   | Vallabhbhai Patel Chest Institute, University of Delhi, Delhi - 110007  |
| सीए अनुमोदन संख्या/CA Approval Number   | dated 24.12.2025  |
| सक्षम प्राधिकारी अनुमोदन तिथि/Competent Authority Approval Date   | 24-12-2025  |
| सक्षम प्राधिकारी द्वारा प्रदान की गई स्वीकृति का संक्षिप्त विवरण/Brief Description of the Approval Granted by Competent Authority | approval obtained for procurement of equipment namely Real Time RT-PCR Machine, Cat # A47326, Name: Quant Studio 5 Dx Real-Time PCR System, Make: ThermoFisher Scientific on proprietary article basis alongwith exemption for MII Purchase Preference as the item is proprietary in nature |

Competent Authority Approval for not opting Make In India Preference : [View Document](#)

#### एमएसई खरीद वरीयता/MSE Purchase Preference

|  |     |
|--|-----|
| एमएसई खरीद वरीयता/MSE Purchase Preference  | Yes |
| सूचना और लघु उधम मूल उपकरण निर्माताओं को खरीद में प्राथमिकता, यदि उनका मूल्य $L1+X\%$ तक की सीमा में हो / Purchase Preference to MSE OEMs available upto price within $L1+X\%$ | 15  |
| सूचना और लघु उधम को खरीद में प्राथमिकता के लिए बिड की मात्रा का अधिकतम प्रतिशत / Maximum Percentage of Bid quantity for MSE purchase preference                                | 25  |

#### मध्यस्थिता खंड के समावेश को मंजूरी देने वाले सक्षम प्राधिकारी का विवरण:/Details of the Competent Authority approval for PAC

Competent Authority Approval document for PAC : [View Document](#)

|   |                 |
|---|-----------------|
| सक्षम प्राधिकारी का नाम/Name of Competent Authority       | Prof. Raj Kumar |
| सक्षम प्राधिकारी का पद/Designation of Competent Authority | Director        |

|   |   |
|---|---|
| सक्षम प्राधिकारी का कार्यालय/विभाग/प्रभाग/Office / Department / Division of Competent Authority                                   | Vallabhbhai Patel Chest Institute, University of Delhi, Delhi - 110007                    |
| सीए अनुमोदन संख्या/CA Approval Number   | dated 24.12.2025  |
| सक्षम प्राधिकारी अनुमोदन तिथि/Competent Authority Approval Date   | 2025-12-24 00:00:00   |
| सक्षम प्राधिकारी द्वारा प्रदान की गई स्वीकृति का संक्षिप्त विवरण/Brief Description of the Approval Granted by Competent Authority | approval obtained for procurement of equipment namely RT PCR on proprietary article basis |

1. If the bidder is a Micro or Small Enterprise (MSE) as per latest orders issued by Ministry of MSME, the bidder shall be relaxed from the eligibility criteria of "Bidder Turnover" as defined above subject to meeting of quality and technical specifications. If the bidder itself is MSE OEM of the offered products, it would be relaxed from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. The bidder seeking Relaxation from Turnover, shall upload the supporting documents to prove his eligibility for Relaxation.
2. Purchase preference will be given to MSEs having valid Udyam Registration and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service and Buyer will decide eligibility for purchase preference based on documentary evidence submitted, while evaluating the bid. If L-1 is not an MSE and MSE Seller (s) has / have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 25% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the OM No. F.1/4/2021-PPD dated 18.05.2023 OM No.1\_4\_2021\_PPD dated 18.05.2023 for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017. Benefits of MSE will be allowed only if seller is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.
3. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.

### Real Time PCR Machine (V2) ( 1 pieces ) ( Under PAC )

|       |                   |
|-------|-------------------|
| Make  | appliedbiosystems |
| Model | a47326            |

### तकनीकी विशिष्टियाँ /Technical Specifications

\* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

| विवरण/Specification | विशिष्टि का नाम /Specification Name | बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values) |
|---------------------|-------------------------------------|---|
| GENERAL INFORMATION | Product Name                        | Real Time PCR Machine   |
| PRODUCT INFORMATION | System Type                         | Open  |
|                     | Capacity of Blocks                  | 96 well   |

| विवरण/Specification    | विवरणीय का नाम /Specification Name   | बिड के लिए आवश्यक अनुमति मूल्य /Bid Requirement (Allowed Values)  |
|------------------------|--|---|
|                        | <b>Number of channels (minimum)</b>  | 2, 3, 4, 5, 6   |
|                        | <b>Maximum Heating ramp rate in degree per second</b>  | 3, 4, 5, 6, 7, 8, 9, 10   |
|                        | <b>Maximum Cooling ramp rate in degree per second</b>  | 3, 4, 5, 6, 7, 8, 9, 10   |
|                        | <b>Adjustable heating / cooling ramp rate</b>  | Yes, No Or higher   |
|                        | <b>Programmable Steps and cycles</b>   | Yes, No   |
|                        | <b>Number of USB ports</b>   | 2, 3, 4, 1  |
|                        | Type of chemistries that can be run on system  | Taqman, SYBR green, Molecular Beacon and all other fluorescent dye based  |
|                        | <b>Compatibility for well strips</b>   | 8, 96   |
|                        | Analysis performed   | Gene Expression, Plus/minus assay, SNP, Allelic Discrimination and Dissociation Curve Analysis, DNA Quantitation, Gene Expression Analysis, Comparative Analysis Curve, Standard Curve/ Relative Standard Curve |
|                        | Suitable software for Data Acquisition, analysis of run and also for Gene Expression analysis by relative quantity or normalized expression  | Yes   |
|                        | <b>Temperature setting accuracy (degree celcius)</b>   | $\pm 0.1, \pm 0.2, \pm 0.3$   |
|                        | <b>Well to well temperature uniformity (degree celcius)</b>  | $\pm 0.2, \pm 0.3, \pm 0.4$   |
| DATA MANAGEMENT SYSTEM | <b>PC hard disk</b>  | 500 GB, 1 TB Or higher  |
| CERTIFICATIONS         | Compliance to Medical Device Rules (MDR) 2017 as amended till date   | Yes   |
|                        | Availability of valid Medical Device license for the product issued from the competent authority defined under Drugs and Cosmetic Act 1940 and Rules made there under as amended till date | Yes   |
| WARRANTY               | <b>Warranty in Years (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)</b>                              | 5 Or higher (year)  |

**Additional Specification Parameters - Real Time PCR Machine (V2) ( 1 pieces )**

| Specification Parameter Name                                     | Bid Requirement (Allowed Values)   |
|--|--|
| Terms and Conditions related to required Technical Specification | 1.Specification provided in GeM Tender document is only for reference and for complete specifications, Bidders are requested to refer the Tender Document uploaded as ATC under the heading (Buyer Added bid specific Terms and conditions) 2. The Technical Evaluation of the bid will be as per the these specification only and the bidder shall quote the equipment and submit technical compliance accordingly. |

\* Bidders offering must also comply with the additional specification parameters mentioned above.

**परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity**

| क्र.सं./S.N<br>o. | परेषिती/रिपोर्टिंग<br>अधिकारी /Consignee<br>Reporting/Officer | पता/Address   | मात्रा /Quantity | डिलीवरी के दिन/Delivery<br>Days |
|-------------------|---|---|------------------|---------------------------------|
| 1                 | Sunil Kumar   | 110007, VALLABHBHAI PATEL<br>CHEST INSTITUTE, NORTH<br>CAMPUS, OPPOSITE SHRI RAM<br>COLLEGE OF COMMERCE,<br>UNIVERSITY OF DELHI | 1                | 60                              |

**Special terms and conditions-Version:1 effective from 16-01-2025 for category Real Time PCR Machine (V2)**

1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules (including Medical Device Rule 2017) made there under as amended till date will always be applicable. This will include all notifications issued by Central Drugs Standard Control Organization (CDSCO), Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
2. The sellers are registered on GeM based on self declaration of valid Medical Device License, product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of Medical Device license, product certification, manufacturer certification/licenses, test reports etc.
3. In case of authorized resellers/distributors, it will be the legal & regulatory liability of the manufacturer to ensure that their resellers/distributors are operating in compliance with all relevant laws and regulations and are properly licensed to sell the manufacturer's products, including verifying the validity and authenticity of Medical Device license held by them.
4. The price offered by the seller/bidder shall not, in any case exceed the DPCO/NPPA controlled price or price fixed by State Government, if any. The seller must reduce the prices if there is any reduction in DPCO/NPPA ceiling price or price fixed by State Government, if any.
5. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions (ATC) in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions (GTC), whenever there are any conflicting provisions.
6. **Comprehensive warranty:** Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables. Further there will be 98% uptime warranty during warranty

period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.

7. **Service centres:** Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address, telephone numbers, e mails etc at time of making the supplies. It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled. Details of toll-free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.
8. **Source of supply:** It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.
9. **Packing and Marking:** Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transhipment (if any), rough handling, open storage etc. without any damage, deterioration etc. The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take into consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date, brief description of goods including quantity, Packing list reference number, country of origin of goods and any other relevant details.
10. **Spare Parts:** Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM. It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies. In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available. OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.
11. **Installation, Training, Manuals:** Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations, training and manuals the same shall also be applicable.
12. **Electrical safety checking:** Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee .They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent .In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.
13. **Software:** All software updates should be provided free of cost during warranty period.

## केता द्वारा जोड़ी गई बिंद की विशेष शर्तें/Buyer Added Bid Specific Terms and Conditions

### 1. Buyer Added Bid Specific ATC

Buyer Added text based ATC clauses

| <b>Sr. No.</b> | <b>Real-Time PCR System</b>   |
|----------------|---|
| 1              | The system should be an automated & integrated system for both real-time PCR and post-PCR (end-point) analysis with the following features: |

| <b>Hardware:</b> |   |
|------------------|---|
| 2                | The system should be a 96-well block system with Peltier-based heating & cooling.   |
| 3                | The excitation source should be LED/Laser and the detection system should be simultaneous and scan-free for all wells with CMOS detection.  |
| 4                | The system should have 6 excitation and 6 emission filters and come along with fixed (scanning-free) optics, ensuring robust data.  |
| 5                | The system should have a temperature range of 4°C - 99.9°C to facilitate all qPCR applications.   |
| 6                | The system should have a peak block ramp rate exceeding 6.5°C/second or more in 0.1 ml block.   |
| 7                | The system should have a better than gradient function in veriflex blocks and six veriflex temperature zones for running 6 different samples with different annealing temperatures simultaneously for different time intervals. |
| 8                | The system should support a reaction volume minimum of 10 - 30 µL in a 0.1ml tube and have more than 5 or 6 color multiplexing without passive reference dye in a single reaction tube.   |
| 9                | The system should be standalone with 10G B onboard memory to save the runs and have an interactive Touch Screen LCD feature.  |
| 10               | Fast-PCR in less than 35 minutes should be an integral feature of the system including collection of up to 21 unique combinations of wavelengths during a single run for multiplexing on the 96-well block instrument.          |
| <b>Software:</b> |   |

|    |   |
|----|---|
| 11 | The instrument should have software that can analyze multiple perspectives in the Multiple Plots view, with side by side views of all data aspects including the amplification plots, standard curve, multicomponent data plots, and raw data. The system should give heat map of the amplification & analyzed data. Software should have PCR efficiency factor correction for gene quantification. |
| 12 | The system should along with software to support applications including absolute quantitation, Relative quantitation, multiplex-PCR, allelic discrimination (SNP), melt curve analysis as well as pathogen detection, plus/minus assay using internal positive control & Melting curve analysis for gene & mutation screening.  |
| 13 | System should be US FDA/European Certification (CE)/ISO/ISI/BIS./IVD approved   |
|    | <b>Supporting Chemistries and Applications:</b>   |
| 14 | The system should be completely open system to support all the Real Time PCR chemistry like TaqMans, SYBR Green, Simple & Hydrolysis Probes, and Molecular Beacons etc.   |
| 15 | The system should detect even 1 copy in a single reaction tube without passive reference dye.   |
| 16 | The system should be open system with flexibility to use micro well plates, individual tubes, and 8-tube strips.  |
| 17 | System should detect differences in target quantity as small as 1.5-fold in single plex reactions and should have 10 logs of linear dynamic range.  |
|    | <b>Terms &amp; Conditions:</b>  |
| 18 | The instrument should come along with 5 years of warranty. Should be supported with remote services, cloud connectivity online monitoring, and external barcode using USB, etc.   |
| 19 | Should be provided with branded laptop for data analysis  |

|    |  |
|----|--|
| 20 | The quoted system must have full license for PCR process and attach a list of minimum 30 installations.  |
| 21 | Firm MUST provide a compliance statement vis-à-vis specifications in a "tabular form" clearly stating the compliance and giving justification, if any supported by technical literature with clear reference of page number, paragraph or lines. This statement must be signed, with the company seal, by the Tendered for its authenticity and acceptance that any incorrect or ambiguous information found submitted will result in disqualification of the Tender |

1. Bidders are requested to fill the tender as per e-procurement notice no. VPCI/PUR/OT/25-26/VRDL/DHR/ICMR/13-18/76 dated (PDF File) attached in buyer added bid specific ATC point no. 2

## 2. Buyer Added Bid Specific ATC

Buyer uploaded ATC document [Click here to view the file.](#)

## अस्वीकरण/Disclaimer

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for attached categories, trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the

case may be.

15. Buyer added ATC Clauses which are in contravention of clauses defined by buyer in system generated bid template as indicated above in the Bid Details section, EMD Detail, ePBG Detail and MII and MSE Purchase Preference sections of the bid, unless otherwise allowed by GeM GTC.
16. In a category based bid, adding additional items, through buyer added additional scope of work/ additional terms and conditions/or any other document. If buyer needs more items along with the main item, the same must be added through bunching category based items or by bunching custom catalogs or bunching a BoQ with the main category based item, the same must not be done through ATC or Scope of Work.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

**All GeM Sellers/Service Providers shall ensure full compliance with all applicable labour laws, including the provisions, rules, schemes and guidelines under the four Labour Codes i.e. the Code on Wages, 2019; the Industrial Relations Code, 2020; the Occupational Safety, Health and Working Conditions Code, 2020; and the Code on Social Security, 2020 as and when notified and brought into force by the Government of India.**

**For all provisions of the Labour Codes that are pending operationalisation through rules, schemes or notifications, the corresponding provisions of the pre-existing labour enactments (such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972, etc. and relevant State Rules) shall continue to remain applicable.**

**The Seller/ Service Providers shall, therefore, be responsible for ensuring compliance under:**

- All notified and enforceable provisions of the new Labour Codes as mentioned hereinabove; and
- All operative provisions of the erstwhile Labour Laws until their complete substitution.

**All obligations relating to wages, social security, safety, working conditions, industrial relations etc. and any other statutory requirements shall be strictly met by the Seller/ Service Provider. Any non-compliance shall constitute a breach of the contract and shall entitle the Buyer to take appropriate action in accordance with the contract and applicable law.**

यह बिड सामान्य शर्तों के अंतर्गत भी शासित है। /This Bid is also governed by the General Terms and Conditions

जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस लियिदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा। In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

**---धन्यवाद/Thank You---**